



Tender enquiry no. ADMIN/Tender/Multipara\_ & Neonatal\_Multipara\_Monitor/1/2015 Date 09-02-2015

## Corrigendum

### **MULTIPARA MONITOR (HIGH END) (14 Nos.) Schedule-A**

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Last date of submission/ opening where mentioned in tender document	Last date of submission is 13.02.2015 at 3:00 PM and opening of technical bid 13.02.2015 at 3:30 PM	Last date of submission is <b>18.02.2015</b> at 3:00 PM and opening of technical bid <b>18.02.2015 at 4:00 PM</b>
1	Page no. 23 Corrigendum dated 06-02-15	<b>Individual Monitors :</b> <ul style="list-style-type: none"> <li>▪ Medical grade, TFT Flat screen, slim size, at least 19" display</li> </ul> <i>(It was deleted on last corrigendum)</i>	<b>Individual Monitors :</b> Medical grade, TFT Flat screen, slim size, at least 19" display
2	Point no. 6 of Corrigendum dated 06-02-15	All parameters in the monitors should be individually USFDA approved.	All parameters in the monitors should be USFDA approved.
3	Point no. 7 of Corrigendum dated 06-02-15	<b>Monitor should have separate transport module with screen and data storage and able to monitor temperature, ECG, SPO2, NIBP &amp; atleast one IBP</b>	<b>Deleted.</b>
4	Page no. 24 Point no. 2	Minimally invasive cardiac output	Minimally invasive cardiac output / Picco Module.
5	Page no. 24 Point no. 5	System must display perfusion index (PI %) from SpO <sub>2</sub> as an indication of pulse strength at the sensor site.	<b>Deleted.</b>
6	Page no. 25 Point no. 8	Monitor must measure dynamic preload parameters such as pulse pressure variation (PPV) and systolic pressure variation (SPV) from arterial waveform and display on main screen.)	Monitor must measure dynamic preload parameters such as pulse pressure variation (PPV) and <b>stroke volume variation (SVV)</b> from arterial waveform and display on main screen.)
7	Page no. 26 Point no. 11	<b>Full waveform disclosure of 24 hrs.</b>	<b>Deleted.</b>

### **NEONATAL MULTIPARA MONITOR -HIGH END(2 Nos.) Schedule-B**

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Point no. 1 of Corrigendum dated 06-02-15	<b>Monitor should have separate transport module with screen and data storage and able to monitor temperature, ECG, SPO2, NIBP &amp; atleast one IBP</b>	<b>Deleted.</b>



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

**MULTIPARA MONITOR (8 Nos. ) Schedule-C**

<b>Sr. No.</b>	<b>Page No/ Clause no/ point no</b>	<b>Existing sentence</b>	<b>To be read as</b>
1	Point no. 1 of Corrigendum dated 06-02-15	<b>Monitor should have separate transport module with screen and data storage and able to monitor temperature, ECG, SPO2, NIBP &amp; atleast one IBP</b>	<b>Deleted.</b>

Store Officer  
AIIMS, Raipur