



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

**Tender enquiry no. ADMIN/Tender/ Tender\_ Syringe\_ Infusion\_ Pumps/ 1/2014**

**Corrigendum for Syringe Infusion Pumps**

**Syringe Infusion Pumps (100 Nos.)**

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	As mentioned in tender document	Last Date of submission 11/10/2014 at 03:00 PM	<b>Last Date of submission 20/10/2014 at 03:00 PM</b>
2	As mentioned in tender document	Technical Bid will be opened on 11/10/2014 at 04:30 PM	Technical Bid will be opened on 20/10/2014 at 04:30 PM
3	Page no. 20	Syringe size: 5, 10, 20, 50ml sizes	Syringe size: <b>5, 10, 20, 30, 50ml</b> sizes
4	Page no. 21	Audible alarm: Sound intensity user-adjustable Rechargeable, maintenance free battery Life when fully charged: up to 10 hrs	<b>Audible alarm: Sound intensity user-adjustable Rechargeable, maintenance free battery Life when fully charged: at least 4 hrs</b>
5	Page no. 23	IP65 or IP55 Body should not allow ingress of liquid. Certificate should be provided.	<b>IP23 Body should not allow ingress of liquid. Certificate should be provided.</b>
6	Page no. 23	Drug Library	<b>Should have Drug Library with 1000 drugs and should have option to change/add/delete drug name</b>
7	Addendum Page no. 20	-	<b>General requirements: Body weight mode</b>

Deputy Director (Admin)  
AIIMS, Raipur