Notice Inviting Tender for

Dentistry Equipment at

All India Institute of Medical Sciences, Raipur

No	DME Stage	Start Date & Time
1.	NIT No.	Admin/Tender/Dentistry_Equipment/1/2014
2.	NIT issue date	02-04-2014
3.	Pre bid Meeting	10-04-2014 at 12:30 PM
4.	Last Date of submission	25-04-2014 at 15:00 PM



All India Institute of Medical Sciences, Raipur Tatibandh, Raipur – 492099, Chhattisgarh

Tele: 0771-2573777, email: <u>admin@aiimsraipur.edu.in</u> Website: www.aiimsraipur.edu.in

<u>Chapter I- Instruction to bidders</u> <u>Notice Inviting Tender</u> "Dentistry Equipment"

All India Institute of Medical Sciences, Raipur, Chhattisgarh, an apex healthcare institute being established by Parliament of India under aegis of Ministry of Health & Family Welfare, Government of India, invites sealed Tenders for supply & installation of the following item(s) at the institute. You are requested to quote your best offer along with the complete detail of specifications, terms & conditions.

Schedule	Name of the Items required to be purchased	Quantity tendered for
А.	Advanced Electronic Dental Chair	02
В.	Table Top Autoclave	01
C.	Dental Digital Radiography System with X Ray Unit	01
D.	Surgical Micromotor With Console	01
E.	LED Light Cure Unit	01
F.	Endomoter - Endodontic Electric Motor	01
G.	Oral Surgery Instruments	Mentioned in Annexure I
H.	General Dental Instruments	Mentioned in Annexure I
I.	Orthodontic Instruments	Mentioned in Annexure I
J.	Endodontic And Restorative Instruments	Mentioned in Annexure I
К.	Periodontics Instruments	Mentioned in Annexure I

- Interested parties may send their tender in sealed cover addressed to the Administrative Officer, All India Institute of Medical Sciences, Tatibandh, Raipur superscripted with tender number and complete in all respects latest by 25-04-2014 up to 3:00 pm. The Quotations will be opened on the same day at 03.30 PM in the Committee Hall, Medical College Building, All India Institute of Medical Sciences, Raipur. The tenders received after the scheduled date and time will be rejected out rightly.
- 2. The tender is in <u>two- bid</u> system i.e. Technical Bid & Financial Bid .The technical bid will be opened on the designated date by the Purchase Committee. The financial bid containing the rate of various item(s) will be opened on a suitable date, to be intimated later by the Purchase Committee in respect of those who qualify the terms and conditions of the technical bid.
- 3. The bidder should seal the technical bid and the financial bid in separate covers superscribed "Technical bid for Tender for Supply of Dentistry Equipment"

and "**Financial Bid for Tender for Supply of Dentistry Equipment**". Both Sealed Envelopes should be kept in a main/ bigger envelope super scribed as "**Tender for Supply of Dentistry Equipment**". The 'Technical Bid" will be analyzed and 'Financial Bid' of only those firms who are found eligible in 'Technical Bid' will be opened in due course and the eligible firms would be intimated there of accordingly.

Issue Date	:	02-04-2014				
Pre bid Meeting	:	10-04-2014 at 12:00 PM at Committee Hall, 2nd floor,				
		Medical Colle	ege Building, AIIMS Raipur			
Venue	:	Deputy Director (Admin) AIIMS, Tatibandh, Raipur-				
		492099				
Last date and time of receipt of tender	:	25-04-2014 at	03:00 PM			
Date & time of opening of tender	:	25-04-2014 at	t 03:30 PM			
Amount of Earnest Money Deposit	:	Schedule A	₹75,000/- (Rupees Seventy Five			
(EMD)			Thousand only)			
		Schedule B	₹5,000/- (Rupees Five Thousand only)			
		Schedule C	₹12,000/- (Rupees Twelve Thousand			
			only)			
		Schedule D	₹6,000/- (Rupees Six Thousand Only)			
		Schedule E	₹2,000/- (Rupees Two Thousand Only)			
		Schedule F	₹2,000/- (Rupees Two Thousand Only)			
		Schedule G	₹15,000/- (Rupees Fifteen Thousand			
			Only)			
		Schedule H	₹5,000/- (Rupees Five Thousand Only)			
		Schedule I	₹2,000/- (Rupees Two Thousand Only)			
		Schedule J	₹ 5000/- (Rupees Five Thousand Only)			
		Schedule K	₹2000/- (Rupees Two Thousand Only)			
Tender Cost	••	₹ 2000/- (Rup	bees two thousand only)			

Schedule of Tender

4. Tender document may be downloaded from this Institute's official website <u>http://www.aiimsraipur.edu.in</u> and the tenderer shall deposit a separate Bank Draft/ Pay Order/ Banker's cheque in favour of "All India Institute of Medical Sciences, Raipur" worth ₹2000/- along with tender Document (Technical Bid). The tenders submitted without tender cost or without EMD shall be liable to be rejected summarily. The cost of the bid document is non-refundable.

Administrative Officer AIIMS, Raipur

Chapter-II- Conditions of Contract

General Terms and Conditions

<u>Subject:-</u> Notice Inviting Tender for Supply of Dentistry Equipment for All India Institute of Medical Sciences, Raipur

1. Earnest Money:

Earnest money by means of a Bank Demand Draft/ Pay Order of ₹75,000/- (Rupees Seventy Five Thousand only) for Schedule A, ₹5,000/- (Rupees Five Thousand only) for Schedule B, ₹12,000/- (Rupees Twelve Thousand only) for Schedule C, ₹ 6,000/- (Rupees Six Thousand Only) for Schedule D, ₹2,000/- (Rupees Two Thousand Only) for Schedule E, ₹ 2,000/- (Rupees Two Thousand Only) for Schedule F, ₹ 15,000/- (Rupees Fifteen Thousand Only) for Schedule G, ₹ 5,000/- (Rupees Five Thousand Only) for Schedule I, ₹ 5,000/- (Rupees Five Thousand Only) for Schedule I, ₹ 5,000/- (Rupees Five Thousand Only) for Schedule I, ₹ 5,000/- (Rupees Five Thousand Only) for Schedule J, and ₹ 2,000/- (Rupees Two Thousand Only) for Schedule I, and ₹ 2,000/-

- a) No request for transfer of any pervious deposit of earnest money or security deposit or payment of any pending bill held by the ministry in respect of any previous work will be entertained.
- b) Tenderer shall not be permitted to withdraw his offer or modify the terms and conditions thereof. In case the tenderer fail to observe and comply with stipulations made herein or backs out after quoting the rates, the aforesaid amount of earnest money will be forfeited.
- c) Tenders without Earnest Money or Tender cost will be summarily rejected.
- d) No claim shall lie against the AIIMS in respect of erosion in the value or interest on the amount of EMD.

2. Preparation and Submission of Tender:

The tender should be submitted in two parts i.e. Technical Bid and Financial Bid. The Technical Bid and the Financial Bid should be sealed by the bidder in two separate covers **"Technical Bid for Tender for Supply of Dentistry Equipment"** and **"Financial Bid for Tender for Supply of Dentistry Equipment"**. Both Sealed Envelopes should be kept in a main/ bigger envelope super scribed as **"Tender for Supply of Dentistry Equipment"**.

3. <u>Signing of Tender:</u>

All the duly filled/completed pages of the tender should be given serial /page number on each page and signed. The individual signing the tender or other documents connected with contract must specify whether he sign as:

- a) A sole proprietor of the concern or constituted attorney of such sole proprietor;
- b) A partner of the firm, if it is a partnership firm in which case he must have authority to execute the contracts on behalf of the firm and to refer to arbitration disputes concerning the business of the partnership either by virtue of the partnership agreement or by a power of attorney duly executed by the partners of the firm.
- c) Director or a principal officer duly authorized by the Board of Directors of the Company, if it is a company.

<u>N.B.</u>

- i. In case of partnership firms, a copy of the partnership agreement, or general power of attorney duly attested by a Notary Public, should be furnished on stamped paper duly sworn or affirmed by all the partners admitting execution of the partnership agreement or the general power of attorney. The attested copy of the certificate of registration of firm should also be enclosed along with the tender.
- ii. In the case of partnerships firm, where no authority to refer disputes concerning the business of partnership firm has been conferred on any partner, the tender and all other related document must be signed by all partners of the firm.
- iii. A person signing the tender form or any document forming part of the tender on behalf of another person should have an authority to bind such other person and if, on enquiry it appears that the person so signing had no authority to do so, AIIMS, Raipur may without prejudice, cancel the contract and hold the signatory liable for all costs, consequences and damages under the civil and criminal remedies available.
- iv. The tenderer should sign and affix his firm's stamp at each page of the tender and all its annexure as the acceptance of the offer made by tenderer will be deemed as a contract and no separate formal contract will be drawn. NO PAGE SHOULD BE REMOVED/ DETACHED FROM THIS NOTICE INVITING TENDER.

4. **Opening of Tender**

The tenderer is at liberty to present either him or authorize not more than one representative to be present at the opening of the tender. The representative present at the opening of the tender on behalf of the tenderer should bring with him a letter of authority from the tenderer and proof of identification

5. <u>Validity of the bids:</u>

The bids shall be valid for a period of 120 day from the date of opening of the tender. This has to be so specified by the tenderer in the commercial bid

6. <u>Right of acceptance:</u>

AIIMS, Raipur reserve the right to accept or reject any or all tenders /quotations without assigning any reason there of and also does not bind itself to accept the lowest quotation or any tender. AIIMS, Raipur also reserves the rights to accept all the equipment/instruments in the given tender or only part of it in any given schedule without assigning any reason. Any failure on the part of the contractor to observe the prescribed procedure and any attempt to canvass for the work will prejudice the contractor's quotation or any tender.

7. <u>Communication of Acceptance / Right of Acceptance:</u>

AIIMS, Raipur, reserves all right to reject any tender including of those tenderers who fails to comply with the instructions without assigning any reason whatsoever and does not bind itself to accept the lowest or any specific tender. The decision of this Institute in this regard will be final and binding. Any failure on the part of the contractor to observer the prescribed procedure and any attempt to canvass the work prejudice the contractor's quotation.

8. <u>Performance Security:</u>

The successful tenderer will be required to furnish a Performance Security Deposit of 10% of tender amount in the form of Fixed Deposit Receipt or Bank Guarantee from any Nationalized Bank duly pledged in the name of the "All India Institute of Medical Sciences, Raipur". The security deposit can be forfeited by order of this Institute in the event of any breach or negligence or non–observance of any condition of contract or for unsatisfactory performance or non–observance of any condition of the contract.

Performance Security shall be submitted in the form of Bank Guarantee or Fixed Deposit Receipt issued by a scheduled Bank and the Performa provided with the tender document. Performance Security will be discharged 60 days after completion of contractor's performance obligations (including warranty period) under the contract.

<u>Sign of Bidder</u>

9. <u>Delivery & Installation:</u>

The Instruments shall be delivered within 50 days of issue of supply order. Satisfactory installation / commissioning and handover of the instruments will be completed within two weeks from the date of receipt of the instruments at the AIIMS, Raipur premises.

10. Inspecting , Testing and Quality control

- i. The purchaser or its nominal representative will inspect and test the ordered goods and the related services to confirm their conformity to the contract specification and other quality control details incorporated.
- ii. The Purchaser reserves the right for stage inspection during manufacturing process, pre-dispatch inspection or post-delivery inspection. The Purchaser as deemed fit shall resort to all these inspection or may waive off any of the inspection.
- iii. Stage Inspection: The Supplier shall offer semi-finished product without paint for stage inspection supported with invoice & test report of major raw material as per relevant specification standards.
- iv. Before resorting to Bulk Manufacture, the supplier shall offer/send the prototype/actual finished sample for approval by the Inspector. The approval of Samples shall be in respect of workmanship and finishing and shall be without prejudice to the rights of purchaser/user to get random testing of items during pre-dispatch inspection or form the actual lot offered. In the event of the failure of the supplier to deliver the sample by the date specified in the contact or any other date agreed by the purchaser or in the event of rejection of the sample second time, the purchaser shall be entitled to cancel the contract and purchase the stores at the risk and cost of the supplier.
- v. The Inspector shall have full and free access at any time during the execution of the contract to the supplier's work for satisfying himself that the goods are being manufactured in accordance with the specification mentioned in the Contract, and he may require the Supplier to make arrangements of inspection of the goods or any part thereof or any material at his premises or at any other place specified by the Inspector and if the Supplier has been permitted to employ the services of a sub-inspection of the goods after they have been inspected by the Inspector for the purpose aforesaid. The decision of the purchaser/ inspection authority in this regard shall be final and binding on the Supplier. All terms and condition of the contract as they apply to the inspection shall also apply to the re-inspection.
- vi. The Supplier shall provide, without any extra charge, all material, tools, labour and assistance of every kind, which the Inspector may demand of him for any test, and examination, which he shall require to be made on the Supplier's

premises and the Supplier shall bear and pay all costs attendant thereon. If the Supplier fails to comply with condition aforesaid, the Inspector shall, in his sole judgment, be entitled to remove for test and examination all or any of the goods manufactured by the Supplier to any premises other than his (Supplier's) and in all such cases the Supplier's shall bear the cost of transport/and carrying out such tests elsewhere. A certificate in writing of the Inspector, that the Supplier has failed to provide the facilities and the means for test and examination, shall be final.

- vii. The Supplier shall also provide and deliver for test, free of charge, at such place other than his premises as the Inspector may specify such material or goods as he may require for tests for which Supplier does not have the facilities or special/independent tests.
- viii. The Inspector shall have the right to put all the goods or materials forming part of the same or any part thereof to such tests as he may think fit and proper. The Supplier shall not be entitled to object on any ground whatsoever to the method of testing adopted by the Inspector.
- ix. Goods accepted by the purchase/AIIMS, Raipur and/or its inspector at initial inspection and in final inspection in terms of the contract shall in no way dilute purchasers/AIIMS, Raipur's right to reject the same later.
- x. The Purchaser's Inspector reserves the right to may draw samples through random sampling method after receiving the finished goods at destination. All the supplies shall be accompanied with the Manufacturer's certificate that the material conforms to the specifications.

Further to above if on the goods being rejected by the Inspector or AIIMS, Raipur at the destination, the Supplier fails to make satisfactory supply within the stipulated period of delivery the Purchaser shall be at liberty to:-

- a. Require the Suppler to replace the rejected goods forthwith but in any event not later than a period of 30 days from the date of rejection and the Supplier shall bear all cost of such replacement, including freight, if any, on such replacing and replaced goods but without being entitled to any extra payment on that or any other account.
- b. Purchase or authorize the purchase of quantity of the goods rejected or goods of a similar description when goods exactly complying with particulars are not in the opinion of the Purchase, which shall be final, readily available without notice to the supplier at his risk and cost and without affecting the Supplier's liability as regards the supply of any further instalment due under the contract, or
- c. Cancel the contract and purchase or authorised the purchase of the goods or goods of a similar description (when goods complying with particulars are not in the opening of the Purchase, which shall be final, readily available) at the risk and cost of the Supplier.

- d. The Inspector shall have the power:-
 - Before any goods or part thereof are submitted for inspection to certify that they cannot be in accordance with the contract owing to the adoption of any unsatisfactory method of manufacture.
 - To reject any goods submitted as not being accordance with particulars.
 - To reject the whole of the installment tendered for inspection, if after inspection of such portion thereof as he may in his discretion thinks fit, he is satisfied that the same is unsatisfactory.
 - To demand all cost incurred in the inspection and/or test. Cost of test shall be assessed at the rate charged by the Laboratory to private persons for similar work however if goods are accepted all cost incurred shall be born by the Purchaser.

Inspector's decision as regards the rejection shall be final and binding on the Supplier.

11. Guarantee / Warranty, Service, Maintenance:

Warranty

The tenderers must quote for 5 years on site warranty from the date of completion of the satisfactory installation as certified by the stipulated committee. The Warranty charges **shall not** be quoted separately otherwise the offer shall be summarily rejected. Also the Bidders should submit their quote for subsequent 5 years on site AMC (without spare parts) / on site CMC (include free labour, repair, other services & spare parts). Failure to comply this condition will entail the rejection of the Bids. The price comparison shall be made taking into account on basic price and post warranty AMC / CMC. The Rate Contracting Authority reserves the right to award AMC / CMC. A.M.C. (without spare parts) shall be quoted for equipments costing upto Rs.5.00 Lacs and C.M.C. (include free labour, repair, other services & spare parts) shall be quoted for equipments costing above Rs.5.00 Lacs. So the price of AMC / CMC should be quoted according to the cost of equipment. The amount of CMC would be released to the supplier on successful completion of the maintainence of that particular year duly certified by the user department.

Service

The supplier will ensure regular maintenance service by the appropriate engineer having the technical know-how of the equipment. The supplier shall also ensure the presence of resident engineer in the geographical location of this city of Raipur so that he attends the call without loss of time.

Maintenance

It will be mandatory for the supplier to enter into a maintenance contract as applicable below:

- a. If the cost of the unit/instrument/equipment is less than INR Five Lakhs, then the supplier has to sign an annual maintenance contract with the purchaser. During this period the supplier shall ensure quality maintenance / check service of minimum 4 times a year in addition to the calls that might be raised by the user department. This contract will be 1 % per annum of the base price (exclusive of taxes / duties) quoted in the financial bid and shall be paid to the supplier at the end of the year with the clearance / satisfaction certificate issued by the user department.
- b. If the cost of the unit/instrument/equipment is more than INR Five Lakhs, then the supplier has to sign a comprehensive maintenance contract with the purchaser. During this period the supplier shall ensure quality maintenance / check service of minimum 4 times a year in addition to the calls that might be raised by the user department. This contract will be 3 % per annum of the base price (exclusive of taxes / duties) quoted in the financial bid and shall be paid to the supplier at the end of the year with the clearance / satisfaction certificate issued by the user department.

Non-compliance of any of these conditions or any of the conditions mentioned in the document renders the purchaser to blacklist the supplier along with appropriate legal actions.

12.Liquidated Damages

If the supplier fails to deliver any or all of the goods or fails to perform the service within the time frame(s) incorporated in the tender, the Purchaser shall, without prejudice to other right and remedies available to the Purchaser under the tender, deduct from the quoted price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of the quoted price. Once maximum 10% of value of delayed supply is reached, Purchaser may consider termination of the tender.

13.FORCE MAJEURE:

If, at any time during the subsistence of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, act of public enemy, civil commotion, sabotage, fire, floods, explosion, epidemics, quarantine restriction, strikers lockout or act of God (hereinafter referred to as events) provided notice of happening of any such eventuality is given by party to other within 21 days from the date of occurrence thereof, neither party hall by reason of such event be entitled to terminate this contract nor shall either party have any claim for damages against other in respect of such non-performance or delay in performance, and deliveries have been so resumed or not shall be final and conclusive.

Further, that if the performance in whole or in part of any obligation under this contract is prevented or delayed by reason of any such event for a period exceeding 60 days, AIIMS, Raipur party may, at least option to terminate the contract.

14.<u>Insolvency etc.:</u>

In the event of the firm being adjudged insolvent or having a receiver appointed for it by a court or any other order under the Insolvency Act made against them or in the case of a company the passing any resolution or making of any order for winding up, whether voluntary or otherwise, or in the event of the firm failing to comply with any of the conditions herein specified AIIMS, Raipur shall have the power to terminate the contract without any prior notice.

15.Breach of Terms and Conditions:

In case of breach of any terms and conditions as mentioned above, the Competent Authority, will have the right to cancel the work order/ job without assigning any reason thereof and nothing will be payable by AIIMS, Raipur. In that event the security deposit shall also stands forfeited.

16.<u>Subletting of Work:</u>

The firm shall not assign or sublet the work/job or any part of it to any other person or party without having first obtained permission in writing of AIIMS, Raipur, which will be at liberty to refuse if thinks fit. The tender is not transferable. One tenderer shall submit only one tender.

17. Right to call upon information regarding status of work:

The AIIMS, Raipur will have the right to call upon information regarding status of work/ job at any point of time.

18.<u>Terms of payment:</u>

Designated committee shall release 100% payment of the total order value after the successful installation of the ordered goods against the submission of the satisfactory installation report.

19.Arbitration

If any difference arises concerning this agreement, its interpretation on payment to the made thereunder, the same shall be settled out by mutual consultation and negotiation. If attempts for conciliation do not yield any result within a period of 30 days, either of the parties may make a request to the other party for submission of the dispute for decision by an Arbitral Tribunal containing Sole Arbitrator to be appointed by the Secretary, Department of Legal Affairs. Such requests shall be accompanied with a panel of names of three persons to act as the sole arbitrator. In case of such arbitrator refusing, unwilling or becoming incapable to act or his mandate having been terminated under law, another arbitrator shall be appointed in the same manner from among the panel of three persons to be submitted by the claimant. The provision of Arbitration and Conciliation Act, 1990 and the rule framed there under and in force

shall be applicable to such proceedings.

20. Legal Jurisdiction:

The agreement shall be deemed to have been concluded in Raipur, Chhattisgarh and all obligations hereunder shall be deemed to be located at Raipur, Chhattisgarh and Court within Raipur, Chhattisgarh will have Jurisdiction to the exclusion of other courts.

Administrative Officer AIIMS, Raipur

Chapter III - Specifications and allied Technical Details

Parameters and Technical Specifications for Executing the Work:

- 1. The participant manufacture/companies having own manufacturing units or their authorized dealer can only participate. Proof of manufacturing is to be attached. Client may also visit and inspect the manufacture set up as deemed fit.
- 2. Manufacturer should be ISO 9001 Certified. Dealers participating should enclose certificate from their parent manufacturer company.
- 3. The firm should be registered and should have the turnover of atleast ₹ 1,00,00,000/- (Rupees One Crore only) of business in India the last three consecutive financial years.
- 4. The firm has to submit a proof of turnover as mentioned above supported by attested documentary materials.
- 5. Tenderer must provide evidence of having supplied to at least 5 reputed government / reputed private organizations in India including at least one government institution.
- 6. AIIMS, Raipur also reserves the rights to accept all the equipment/instruments in the given tender/schedule or only part of it in any given tender/schedule without assigning any reason.
- 7. The delivery of the items will have to be made at AIIMS, Raipur. No transportation/ cartridge charges will be provided for the same.
- 8. The rate quoted should be firm and final and written in ink or typed against each item and should in no case be overwritten.
- I / We hereby accept the terms and Conditions given in the tender

(Signature & Stamp of the bidder)

Note- Please sign each page of document including terms & conditions & tender

Chapter - IV Contract Form

TENDER FORM - 1 - TECHNICAL INFORMATION AND UNDERTAKING.

(Tenderer may use separate sheet wherever required)

(In separate sealed Cover-I super scribed as "Technical Bid")

		Document Required	Document Supplied (Yes/No)	If yes, provide Reference page number in techno- commercial bid
1.	Name & Address of the manufacture and their authorised dealers/ distributors/Agency with phone number, email, name and telephone/mobile			
2.	Specify your firm/company is a manufactures/ authorised dealer/distributor/ Agency			
3.	Name, Address & designation of the authorized person (Sole proprietor/partner /Director)	Relevant document		
Α	Original Equipment Manufacturer's (OEM) Prequalification:			
	 (a) Please provide the name, address, and the other details of the OEM. If the OEM is originally incorporated in other country, please furnish such details for India also. (b) Please provide full details of factory /manufacturing units for each item for which the bid is invited. 			
4.	OEM must be a ISO 9001 Certified company consistently	Copy of Valid certificates		
5.	OEM must be present in India for at least 5 Years with business history	Company Incorporation certificate		
6.	OEM must have 10+ years of experience supplying and servicing products	Complete installation reports which are older than 2 years		
В	About the tender Offer			
7.	Tenderer must provide evidence of having supplied to at least 5 reputed government / reputed private organizations in India including at least one government institution.	 Copy of the all the Supply orders. Satisfaction certificates (along with contact details) from atleast two institutions. 		

C	Tenderer's prequalification		
8.	Tenderers, if not OEM, must have presence in India for last 10 years and have a turnover of at least Rs. 1 crore for each of the last 3 years	 Company Incorporation Certificate, Copies of authenticated balance sheet for the past three financial years 	
9.	Tenderers, if not OEM, must submit OEM's authorization certificate that the tenderer is authorized for selling, and maintaining the equipment quoted for.	OEM's authorization letter	
10.	* * * *	Undertaking from OEM	
11.	Tenderers must submit letter from OEM mentioning to agree to provide all post- sale installation and maintenance support	Undertaking from OEM	
D	Tender Related		
D 12.			
	(EMD)		
	A. Schedule A - ₹ 75,000/-	FDR/DD No: Date: Payable at	
	B. Schedule B - ₹ 5,000/-	FDR/DD No: Date: Payable at	
	C. Schedule C - ₹ 12,000/-	FDR/DD No: Date: Payable at	
	D. Schedule D - ₹ 6,000/-	FDR/DD No: Date: Payable at	
	E. Schedule E - ₹2,000/-	FDR/DD No: Date: Payable at	
	F. Schedule F - ₹ 2,000/-	FDR/DD No: Date: Payable at	
·	G. Schedule G - ₹ 15,000/-	FDR/DD No: Date: Payable at	
	H. Schedule H - ₹ 5,000/-	FDR/DD No: Date: Payable at	
	I. Schedule I - ₹ 2,000/-	FDR/DD No: Date: Payable at	
ŀ	J. Schedule J - ₹ 5000/-	FDR/DD No:	

			1	
		Date:		
	K. Schedule K - ₹2000/-	Payable at		
	K. Schedule K - K 2000/-	FDR/DD No: Date:		
		Payable at		
13.	Details of the cost of the Tender			
	documents worth ₹ 2000/-(Rupees Two			
	thousand only)			
14	Whether each page of NIT and its annexure			
	have been signed and stamped			
15	Whether Bidders have quoted for each and			
10.	every item mentioned in respective			
	schedule			
	i. Schedule A			
	ii. Schedule B			
	iii. Schedule C			
	iv. Schedule D			
	v. Schedule E			
	vi. Schedule F			
	vii. Schedule G			
	viii. Schedule H			
	ix. Schedule I			
	x. Schedule J			
	xi. Schedule K			
E	Other Relevant Information required			
16.	Permanent Account Number	Copy of the PAN Card		
17.	TIN No. With Proof			
18.	Please attach copy of last of Income Tax			
	Return			
19.	VAT/Service Tax Registration Number.	Please attach copy		
20.	Acceptance of terms & conditions attached	Please sign each		
	(Yes/No).	page of terms and		
		conditions as token		
		of acceptance and		
		submit as part of		
		tender document		
		with technical bid.		
		Otherwise your		
		tender will be		
		rejected.		
21.	Please submit a notarised affidavit on			
	Indian Non judicial stamp paper of Rs. 10/-			
	that no case is pending with the police			
	against the Proprietor/firm/partner or			
	the Company (Agency). Indicate any			

	convictions in the past against the Company/firm/partner. Please also declare that proprietor/firm has never been black listed by any organization.	
22.		
23.	Kindly mention the total number of pages in the tender document	

Note:

- i. Page number/serial number may be given to each and every page of Tender Documents and photocopies of the attested documents attached. Mention Page number, wherever the copy (ies) of the document(s) is kept.
- ii. In case of non-fulfillment of any of the above information/ document(s), the Tender will be summarily rejected without giving any notice.

(Dated Signature of the Tenderer with stamp of firm)

Declaration by the Tenderer:

- 1) That I/we have carefully studied all the terms & conditions of NIT and shall abide by it.
- 2) That I/We shall supply the items of requisite quality.
- 3) That I/We undertake that the information given in this tender are true and correct in all respect and I/We hold the responsibility for the same.

(Signature of Tenderer with seal)

Name:

Address :

"Dentistry Equipment" AIIMS, Raipur <u>FINANCIAL BID for SCHEDULE – A/B/C/D/E/F</u>

S. No.	Item Description	Unit Pric	e in INR	Custom INR (if ap	duty in oplicable)	Taxes (if applicable) VAT / Sales Tax/etc in INR		applicable) VAT / Sales Tax/etc in		Service applicable		Qty of units	in INR if	r Charges applicable ccify)	Total in I price x Q Other C	
		Figures	Words	Figures	Words	Figures	Words	Figures	Words		Figures	Words	Figures	Words		
1	Equipment/ Instrument															
2																
3	Accessories															
	AMC/CMC Total for Five Years (Details is to be mentioned in separate Table)	Grand To	tal*:			•										

(In sealed Cover-II super scribed)

The grand total * will be the deciding factor for L1 provided all the other conditions mentioned in the tender document (GCC, SCC and the technical conditions) are fulfilled.

DATE:

SIGNATURE WITH NAME AND SEAL

"Dentistry Equipment" AIIMS, Raipur <u>FINANCIAL BID for SCHEDULE – G/H/I/J/K</u>

(In sealed Cover-II super scribed "Financial Bid")

S. No.	Item Description	Unit Pric	e in INR	Custom INR (if ap		Taxe applicabl Sales Ta IN	e) VAT / x/etc in	Service applicable	,	Qty of units	in INR if	r Charges applicable cify)	Total in I price x Q Other C	harges)
		Figures	Words	Figures	Words	Figures	Words	Figures	Words		Figures	Words	Figures	Words
1.														
2.														
3.														
4.														
5.														
6.														
7.														
8.														
9.														
10.														
Years	CMC Total for Five (Details is to be mentioned te next Table)		I	<u>.</u>	<u>.</u>	<u>.</u>	I	<u>.</u>	1	I			rand Total	

The Cost of individual equipment/instrument will be the deciding factor for L1 provided all the other conditions mentioned in the tender document (GCC, SCC and the technical conditions) are fulfilled.

DATE:

SIGNATURE WITH NAME AND SEAL

<u>PRICE SCHEDULE FOR ANNUAL MAINTENANCE CONTRACT</u> <u>(A.M.C.) / COMPREHENSIVE MAINTENANCE CONTRACT (C.M.C) AFTER</u> <u>EXPIRY OF WARRANTY</u>

(RATES SHOULD BE QUOTED IN INDIAN RUPEES ONLY)

Sr	SM	Name of	For Sixth	For	For	For Ninth	For Tenth
No	Е	the	year	Seventh	Eighth	year with	year with
	Со	Equipment	with	year with	year with	spare	spare
	de		spare	spare	spare	parts &	parts &
	No.		parts &	parts &	parts &	labour	labour
			labour	labour	labour		
(1)	(2)	(3)	(4)	(5)	(6)	(7)	(8)

Place: Date:

> Signature Name in Capital Letters Designation

SPECIAL CONDITIONS

- 1. The bidders should submit the relevant certificates / evidence of previous supply to other hospitals. If required, the technical committee may enquire from the other hospitals where the bidders have supplied the material.
- 2. The bidders should also submit a certificate from the relevant authority as to the quality of the equipment.
- 3. The bidder should not have been blacklisted before.
- 4. The bidders shall also arrange for the demonstration of their equipment to the concerned committee regarding the quality aspect.

The above quote should include all applicable taxes. If the rates of various items are L1 for different Tenderer, the AIIMS Raipur reserve the right to either accept the L1 of different firm/agencies or will negotiate, with the firm who has quoted the maximum gross value of L1 items to lower the rate of other item up to the limit of L1 quoted by other firms. In this context, final decision of the committee will be binding to all and no claim in the regard can be entertained. The quantity indicated is tentative and may vary, and any decision in this regard by Director AIIMS Raipur shall be final.

Warranty, Service, Maintenance:

Warranty

The equipment price should be quoted along with full comprehensive warranty of 5 years with proper maintenance service mentioned in this document.

Service

The supplier will ensure regular maintenance service by the appropriate engineer having the technical know-how of the equipment. The supplier shall also ensure the presence of resident engineer in the geographical location of this city of Raipur so that he attends the call without loss of time.

Maintenance

It will be mandatory for the supplier to enter into a maintenance contract as applicable below:

- a. If the cost of the unit/instrument/equipment is less than INR Five Lakhs, then the supplier has to sign an annual maintenance contract with the purchaser. During this period the supplier shall ensure quality maintenance / check service of minimum 4 times a year in addition to the calls that might be raised by the user department. This contract will be 1 % per annum of the base price (exclusive of taxes / duties) quoted in the financial bid and shall be paid to the supplier at the end of the year with the clearance / satisfaction certificate issued by the user department.
- b. If the cost of the unit/instrument/equipment is more than INR Five Lakhs, then the supplier has to sign a comprehensive maintenance contract with the purchaser. During this period the supplier shall ensure quality maintenance / check service of minimum 4 times a year in addition to the calls that might be raised by the user department. This contract will be 3 % per annum of the base price (exclusive of taxes / duties) quoted in the financial bid and shall be paid to the supplier at the end of the year with the clearance / satisfaction certificate issued by the user department.

Non-compliance of any of these conditions or any of the conditions mentioned in the document renders the purchaser to blacklist the supplier along with appropriate legal actions.

Declaration by the Bidder:

- 1. This is to certify that I/We before signing this tender have read and fully understood all the terms and conditions contained in Tender document regarding terms & condition of the contract, rules regarding purchase of Maxillofacial Surgery Instruments . I/we agree to abide them.
- 2. No other charges would be payable by Client and there would be no increase in rates during the Contract period.

Place:.....

(Signature of Bidder with seal)

Name : Seal : Address:

Annexure-I

TECHNICAL SPECIFICATION FOR DENTISTRY EQUIPMENT

TECHNICAL SPECIFICATIONS

General Specifications:

- 1. The instruments quoted should be of high quality and ASTM F899 or of equivalent standard specifications for the material used
- 2. All the material /equipment should be European CE/ US FDA/ EN ISO Certified apart from their individual prerequisite mentioned against the material.
- 3. All the Electronic equipment should comply with Electrical safety conforms to standards for electrical safety IEC 60601-1.
- 4. The radiological equipment should be AERB Type Approved
- 5. All the equipment's power input should be 220- 240 V AC, 50Hz fitted with Indian plug.

SCHEDULE –A: ADVANCE ELECTRONIC DENTAL CHAIR

Model Quoted:....

S No	Specifications	Compliance by the Bidder
A1	Fully motorized, pneumatically / electrically	
	driven, which gives smooth and non-jerky start	
	and stop.	
A2	Lowest height range should be between 300 -	
	450 mm to improve visibility and access.	
A3	Chair should have safety brake system while	
	going down for patient exit position	
A4	Chair should have toe movement. While	
	backrest moves down, toe should move up	
A5	The design should enable the operator to be	
	close to the patient to provide optimum vision	
	of the operating field and safe control of all	
	component devices	
A6	The base and other structure should have a	
	corrosion resistant coating	
A7	The backrest should be ultra thin, flexible,	
	highly comfortable, seamless long life	
	upholstery and should disinfectible	
A8	The chair should be designed to provide good	
	ergonomics for both operator and assistant	
A10	Chair should have adjustable ergonomic	
	headrest	
A11	Should be Noiseless DC Motor	

		<u></u>
A12	The chair movement control should be at both	
	finger tip panel and user friendly foot control	
110	with all the functions	
A13	Should have integrated power supply for fiber optic handpieces, piezo electric motor etc	
A14	All the outlet & inlet for the services to the	
Шт	chair should be concealed in the box to be at	
	the foot area of the chair or within the unit, as	
	an infection control measure	
Denta	al Operatory unit	
A15	Should be over head delivery system with	
	minimum 5 delivery ports for various	
	handpieces, scaler and 3-way syringe (tip	
	autoclavable, with 6 spare tips)	
A16	Should have minimum two high Speed Air	
	Rotor terminals with water control on coupling	
	with at least one fiber optic terminal	
A17	It should have one in-built Piezon LED	
	Ultrasonic Scaler (frequency 28-36 KHz) with 4	
	scaler tips and one set of perio-curette tips	
A18	The over head delivery system should have	
	balanced flax arms with pneumatic bracket	
A19	The handpiece control block should flow	
	through water design to eliminate stagnant	
	water	
A20	Built-in antiretraction valves and flush	
	valve system for infection control	
A21	Autoclavable quick disconnect water syringe	
A22	Should have one Air Motor terminal with speed	
	control preferably fiber optic, with speed range	
	of 300 - 40000 RPM in standard mode with	
	cutting power in the range between 50 - 70	
	watt. Should be provided with the Brushless	
	Micromotor cord.	
A23	The assistant side should have high vacuum	
	suction, Medium vacuum suction and 3 way	
	syringe. The Suction system should be compact,	
	unique universal semi wet suction. Durr	
	Suction is preferably.	
A24	X-Ray viewer (LED Based)	
	•	
A25	It should be supplied with following handpieces (all European CE Certified Genuine instruments)	

r		[]
	1. Fiber optic Air rotor nos. 1. (scratch	
	resistant, Titanium Body)	
	2. Fiber optic Micromotor straight nos. 1	
	(scratch resistant, Titanium Body)	
	3. Micro motor contra angle miniature	
	nos. 1,	
A26	Dental Camera system with >15" LCD Monitor	
Cuspi		
A27	Movable cuspidor box and Movable assistant	
	control system should have	
	1. Saliva ejector	
	2. Autoclavable High volume evacuator	
	3. Autoclavable 3-way syringe	
	4. High quality stain proof vitreous China	
	bowl with adjustable cup fill and bowl	
	rinse timers	
0.000	5. Clean water bottle system	
A28	Ating Light	
AZO	Should have latest Sensor operated Non Touch (On/Off) LED Light	
A29	With luminosity of 20000 lux and 25000 lux	
AL J	with maximum degrees of rotation of light arm	
	movements	
A30	Light Head with axial movements - Horizontal,	
1150	Vertical, Axial and diagonal adjustment	
A31	LED light 5000 K cool light or similar high	
	quality light	
A32	Three position intensity with high, medium and	
	composite setting	
Docto	pr's stool	
A33	Cast-metal/alloy base with five tile castors	
A34	Two way adjustable lumbar support	
A35	Integral Gas cylinder for height adjustment	
A36	Height range between 400 - 700 mm	
	tant's stool	
A37	Cast-metal/alloy base with five tile castors	
A38	Height adjustable torso support with height	
	adjustable foot ring	
A39	Integral Gas cylinder for height adjustment	
A40	Height range between 500 - 800 mm	
-	ressor:	
A41	A suitable Medical grade oil Free Compressor,	
	which should be Noise less, and Minimum of	
	1HP, with Retraction Valve and Pressure quage.	
	Noise: 64db maximum	
	(Only one compressor required for both the	
	chairs)	

Prere	equisite	
A42	Demonstration of the Quoted item is must at	
	the predetermined place by the purchaser	
Envir	onmental Factors	
A43	The unit should be capable of being stored	
	continuously in ambient temperatures of $0 - 50$	
	degree centigrade and relative humidity of 15 –	
	90 %.	
A44	The unit should be capable of operating	
	continuously in ambient temperatures of 10 –	
	45 degree centigrade and relative humidity of	
	15 – 90 %.	
Pow	er Supply	
A45	Power input to be 220 – 240V AC, 50 Hz	
Certi	fication	
A46	US –FDA and CE - European approved	
A47	ISO approved	
A48	Manufacturer should have ISO certification for	
-	quality standards	
A49	Electrical safety conforms to standards for	
	electrical safety IEC 60601-1 (OR EQUIVALENT	
	international/national standard) General	
	requirement for Electrical safety of Medical	
	Equipment.	
Syste	m Configuration Accessories, spares and consuma	bles
A50	System as specified on turn key basis	
A51	All consumables required for installation and	
	standardization of system to be given free of cost.	
A52	Accessories as specified	
A53	Provision for modular furniture with sink for	
	dental operator 12 feet x 12 feet should be	
	included in quote.	
A54	Necessary cables and all other accessories those	
	are required for the smooth functioning of the	
	system to be supplied	
	mentation	
A55	User/Technical/Maintenance manuals to be	
	supplied in English	
A56	Certificate of calibration and inspection.	
A57	List of important spare parts and accessories	
	with their part number and costing	
A58	List of Equipment available for providing	
	calibration and routine Preventive	
	Maintenance Support. As per manufacturer	
	documentation in service/technical manual.	
A59	Log book with instructions for daily, weekly,	
		1
	monthly and quarterly maintenance checklist. The job descriptin of the hospital technician	

	1	
	and company service engineer should be	
	clearly spelt out.	
Turn	key: The supplier or his representative shall inspe	ct the site before bidding. The
turnk	ey price should be quoted separately.	
A60	All the necessary civil /electrical /Plumbing	
	other changes required for the effective	
	functioning of the Dental Chair would be done	
	by the supplier.	
A61	Interior furnishing with the requisite furniture,	
	storage, cupboards, etc. for the Operatory area	
	Size of 12 X 12 per dental Chair. Will also form	
	an essential component of this turnkey project.	
A62	Cabinetry, drawers can be opened by push-	
	latch or with handles. With clean lines and a	
	large storage capacity with provision for	
	Proper stacking of Instruments.	
A63	The fixtures should include modular work	
	station with a size of 6 x 2.5 ft. with acid and	
	fire resistant table top, high quality wash basin	

SCHEDULE – B: TABLE TOP AUTOCLAVE

Model Quoted:....

S No	Specifications	Compliance by the Bidder
B1	Sterilizer Type: Table Top Sterilizer.	
B2	Capacity: minimum 18 liters	
B3	Chamber Size: The sterilizer should have	
	Rectangular/Cylindrical chamber with suiting the	
	volume for	
B4	Maximum processing capacity per charge.	
B5	Quality System Compliance: Sterilizer should	
	comply the quality systems as per ISO	
	9001:2000, EN ISO	
B6	Quality Assurance: Sterilizer should be CE or	
	FDA or BIS Certified	
B7	Types of Cycles Process: Table Top Sterilizers	
	should be equipped with B- process as per latest	
B8	International standards	
Cham	ber:	
B9	Should be made of S.S.316 & should comply the	
	Pressure Equipment Directive (PED) & EN 13445	
	norms or equivalent	
B10	Chamber should have minimum 5 years warranty	

	or should confirm 11 50,000 process minimum	
	or should confirm 44- 50,000 process minimum	
D11	life.	
B11	Chamber should have working pressure 2.2 bar &	
D10	design pressure upto 3.8 bar.	
B12	Chamber should have Stress & Fatigue analysis	
	reports for material & construction of the	
	pressure vessel.	
B13	Chamber should be equipped with electrically	
	heated jacket for preheating on stand by mode	
Design		1
B14	Should have horizontal sliding/Hinged door with	
	silicon elastomer rubber gasket to withstand	
	temperature upto 140°C.	
Air Fil	ter	
B15	A disposable air filter should be provided for	
	filtering the atmospheric air before entering	
	inside the chamber. The filter separation	
	efficiency should be higher than 99.998% for	
	particle size less than 0.3μ m	
Cycle]	programs:	
B16	134°C Wrapped & Unwrapped	
B17	121°C Wrapped & Unwrapped	
B18	134°C Flash/Rapid open instrument cycle.	
B19	134°C Textile.	
B20	134°C Prion.	
B21	Test programs: Bowie & Dick, Leak Test, Helix	
	Test	
Water	Storage Tank	
B22	: Sterilizer should have inbuilt water reservoir with	
	storage capacity up to 5 Litres. The water reservoir	
	should have easy access for cleaning & to avoid bio	
	film.	
Steam	Generator	
B23	Sterilizer should have inbuilt steam generator	
	with warranty of 5 years on heating elements.	
	The steam generator design should be with	
	integrated energy storing system for building up	
	power for sterilization loads in short time	
Contro	ol Panel:	1
B24	The control system should be microprocessor	
	based PLC system specially designed for	
	sterilization applications. The control system	
	stermzation applications. The control system	

	should have CPU processor with battery	
	back- up, Digital input/output controls, analog	
	measuring inputs & COM ports for printer & PC	
	connectivity.	
	Alarms	
B25	Automatic process checking & failure correction	
	should be possible by the control system. The	
	range of alarm should include Temperature &	
	pressure sensor failure, phase time- out, doors	
	not properly closed, power failure (less than 10	
	sec should be ignored), continuous self checking	
	of all the safety devices, low water level etc. All	
	the alarms should be audio- visual	
	Accessories	
B26	The sterilizer unit should included Rack with 3	
	levels & suitable size instrument trays should be	
	the part of the supply for every sterilizer. The	
	Sterilizer should have water circulation system	
	so that no drain point & fixed water inlets	
	required	
	Standards & Norms	
B27	The sterilizer must comply the following	
	standards, ISO 9001:2000 (Quality Systems), ISO	
	13485:2003 (Quality Systems for Medical	
	Devices), ISO 14001 (Environment Management	
	System).	
B28	The product should be US FDA and European CE	
	Certified, BIS/ISI approved	
B29	Demonstration of the Quoted item is must at the	
		1

SCHEDULE – C: DENTAL DIGITAL RADIOGRAPHY SYSTEM WITH X RAY UNIT

Model Quoted	:
X ray unit Model Quoted	:
AERB type approval	:

C1 C2 C3 C4	Should be an AERB type approved for the quoted model. Valid approval certificate to be enclosed. High resolution Digital sensor based on CCD/CMOS technology Thickness of the sensor should be less than or equal to 5 mm with Minimum active area - 800 mm area	
C3	CCD/CMOS technology Thickness of the sensor should be less than or equal to 5 mm with Minimum active area - 800	
	equal to 5 mm with Minimum active area - 800	
C4		
	Spatial resolution approx. 25 line pairs/mm	
C5	Cable length for the sensor should be a minimum of 2.5 meters.	
C6	Dynamic range (accurate measurement of bone density), should be more than or equal to 14 bit mage acquisition. (provide data sheet)	
C7	Should have positioning devices i. Bitewing ii. Periapical iii. Endodontic	
C8	Should have at least 10,000 numbers of disposable plastic sheaths for the placement of sensors in the patient's mouth	
C9	Software should be user friendly, accurate, Dental Implant friendly	
Intra o	oral X - ray unit	
C10	AERB Type Approved	
C11	Should be based on DC current	
C12	Tube voltage, selection: 60 - 65 - 70 kVp,	
C13	Tube current 6 mA/ 8 mA,	
C14	Focal spot 0.8 x 0.8 mm	
C15	Total filtration > 2 mm Al	
C16	Minimum range of exposure time range - 0.02 to 3.2 secs,	
C17	Manufactured with international Safety standards for radiation leakage	
C18	Electronic selection of exposure time/radiation according to tooth number.	
C19	It should be possible to select exposure time manually	
C20	Option for wall mount/ mobile stand	

apron C22 Reusable Digital X-ray holders Power Supply C23 Power input to be 220-240VAC, 50Hz fitted with Indian plug C24 Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz) Standards, Safety and Training C25 The product should be US FDA and European CE Certified. C26 Should Be AERB Type Approved C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 User/Technical/Maintenance manuals to be supplied in English with complete data sheet. C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite	Syste	m Configuration Accessories, spares and consuma	ables
C22 Reusable Digital X-ray holders Power Supply C23 Power input to be 220-240VAC, 50Hz fitted with Indian plug C24 Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz) Standards, Safety and Training C25 The product should be US FDA and European CE Certified. C26 Should Be AERB Type Approved C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 User/Technical/Maintenance manuals to be supplied in English with complete data sheet. C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the	C21	X-ray unit should be supplied with lead free	
Power Supply C23 Power input to be 220-240VAC, 50Hz fitted with Indian plug C24 Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz) Standards, Safety and Training C25 The product should be US FDA and European CE Certified. C26 Should Be AERB Type Approved C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 User/Technical/Maintenance manuals to be supplied in English with complete data sheet. C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the		apron	
C23 Power input to be 220-240VAC, 50Hz fitted with Indian plug C24 Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz) Standards, Safety and Training C25 The product should be US FDA and European CE Certified. C26 Should Be AERB Type Approved C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the	C22	Reusable Digital X-ray holders	
Indian plugC24Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz)Stand=rds, Safety and TrainingC25The product should be US FDA and European CE Certified.C26Should Be AERB Type ApprovedC27Manufacturer/ Supplier should have ISO certification for quality standards.C28Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450DocumentationC29User/Technical/Maintenance manuals to be supplied in English with complete data sheet.C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerequisiteC32Demonstration of the Quoted item is must at the	Powe	r Supply	
C24 Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz) Standards, Safety and Training C25 The product should be US FDA and European CE Certified. C26 Should Be AERB Type Approved C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the	C23	Power input to be 220-240VAC, 50Hz fitted with	
meeting ISI Specifications. [Input 160-260 V and output 220-240 V and 50 Hz] Standards, Safety and Training C25 The product should be US FDA and European CE Certified. C26 Should Be AERB Type Approved C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 User/Technical/Maintenance manuals to be supplied in English with complete data sheet. C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32		Indian plug	
output 220-240 V and 50 Hz) Stand=rds, Safety and Training C25 The product should be US FDA and European CE Certified. C26 Should Be AERB Type Approved C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 User/Technical/Maintenance manuals to be supplied in English with complete data sheet. C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the	C24	Servo Voltage stabilizer of appropriate ratings	
Standards, Safety and Training C25 The product should be US FDA and European CE Certified. C26 Should Be AERB Type Approved C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 User/Technical/Maintenance manuals to be supplied in English with complete data sheet. C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the		meeting ISI Specifications.(Input 160-260 V and	
C25 The product should be US FDA and European CE Certified. C26 Should Be AERB Type Approved C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 User/Technical/Maintenance manuals to be supplied in English with complete data sheet. C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the		output 220-240 V and 50 Hz)	
Certified.C26Should Be AERB Type ApprovedC27Manufacturer/ Supplier should have ISO certification for quality standards.C28Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450DocumentationC29User/Technical/Maintenance manuals to be supplied in English with complete data sheet.C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerequisiteC32Demonstration of the Quoted item is must at the	Stand	ards, Safety and Training	
C26Should Be AERB Type ApprovedC27Manufacturer/ Supplier should have ISO certification for quality standards.C28Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450DocumentationC29User/Technical/Maintenance manuals to be supplied in English with complete data sheet.C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerequisiteC32Demonstration of the Quoted item is must at the	C25	The product should be US FDA and European CE	
C27 Manufacturer/ Supplier should have ISO certification for quality standards. C28 Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450 Documentation C29 User/Technical/Maintenance manuals to be supplied in English with complete data sheet. C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the		Certified.	
certification for quality standards.C28Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450DocumentationC29User/Technical/Maintenance manuals to be supplied in English with complete data sheet.C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerequisiteC32Demonstration of the Quoted item is must at the	C26	Should Be AERB Type Approved	
C28Electrical safety for dental x-ray unit conforms to standards for electrical safety IEC-60601 / IS-13450DocumentationC29User/Technical/Maintenance manuals to be supplied in English with complete data sheet.C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerequisiteC32Demonstration of the Quoted item is must at the	C27	Manufacturer/ Supplier should have ISO	
standards for electrical safety IEC-60601 / IS-13450IS-13450DocumentationC29User/Technical/Maintenance manuals to be supplied in English with complete data sheet.C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerewisiteC32Demonstration of the Quoted item is must at the		certification for quality standards.	
IS-13450DocumentationC29User/Technical/Maintenance manuals to be supplied in English with complete data sheet.C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerewisiteC32Demonstration of the Quoted item is must at the	C28	Electrical safety for dental x-ray unit conforms to	
Documentation C29 User/Technical/Maintenance manuals to be supplied in English with complete data sheet. C30 List of important spare parts and accessories with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the		standards for electrical safety IEC-60601 /	
C29User/Technical/Maintenance manuals to be supplied in English with complete data sheet.C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerequisiteC32Demonstration of the Quoted item is must at the		IS-13450	
supplied in English with complete data sheet.C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerequisiteC32Demonstration of the Quoted item is must at the	Docu	mentation	
C30List of important spare parts and accessories with their part number and costingC31Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.PrerequisiteC32Demonstration of the Quoted item is must at the	C29	User/Technical/Maintenance manuals to be	
with their part number and costing C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the		supplied in English with complete data sheet.	
C31 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the	C30	List of important spare parts and accessories	
monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the		with their part number and costing	
The job description of the hospital technician and company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the	C31	Log book with instructions for daily, weekly,	
company service engineer should be clearly spelt out. Prerequisite C32 Demonstration of the Quoted item is must at the		monthly and quarterly maintenance checklist.	
out.PrerequisiteC32Demonstration of the Quoted item is must at the		The job description of the hospital technician and	
Prerequisite C32 Demonstration of the Quoted item is must at the		company service engineer should be clearly spelt	
C32 Demonstration of the Quoted item is must at the			
C C	Prere	quisite	
predetermined place by the purchaser	C32	-	
		predetermined place by the purchaser	

SCHEDULE – D: SURGICAL MICROMOTOR WITH CONSOLE

Model Quoted

:....

S No	Specifications	Compliance by the Bidder
D1	US FDA and European CE Certified	
D2	Should be a Motor-Brushless induction motor	
D3	The Control Unit should have easy control of	
	speed, motor direction, torque and irrigation	
	pump.	
D4	It should work with in 50-60 Hz or 100-240V	
D5	The irrigation pumpage should not be more than 120ml/minute.	
D6	The Max rpm should be 45000 u/min	
D7	Motors should have high consistent torque	
D8	The micromotor should be with a illumination (LED Based/Fiber optic)	
D9	The MicroMotor with Cable should be autoclavable upto 134°C.	
D10	The MicroMotor should come with an intra coupling ("E" coupler) so that allows for universal attachment between motors and handpieces regardless of make	
D11	The Micromotor with Intra coupling should be upgradeable such that Microcompass saw and Micro sagittal saw can be attached in future	
D12	The Control unit should also come with a footpedal	
D13	The Footpedal should have protection proof class of atleast IP x 5.	
D14	The Footpedal should have on and off for irrigation pump and have change over function for clock wise and anticlockwise rotation.	
D15	Should include Cleaning & Lubricating spray	
D16	All the accessories for basic functions should be provided.	
D17	One straight Handpiece. (Should be able to accept regular 701/702 Burs)	
D18	One Contra Angle Hand Piece (Should be able to accept regular 701/702 Burs)	
D19	Demonstration of the Quoted item is must at the predetermined place by the purchaser	

SCHEDULE – E: LED LIGHT CURE UNIT

Model Quoted

:....

S No	Specifications	Compliance by the Bidder
E1	US FDA / European CE Certified	
E2	Should be Cordless	
E3	Ultra-quiet, ultra-portable, and ultra-powerful LED	
	dental curing light.	
E4	Fan free/micro fan silent operation	
E5	Multiple setting light timer	
E6	Light source: Ultra high intensity light emitting	
	diode (LED) 390-510 mm wave length	
E7	The LED Tip Should be freely adjustable, rotating	
	360, for easy access to the restoration site.	
E8	Should include Eye Protection Shields,	
E9	Should have Built-In Radiometer	
E10	Should be slim and not a robust design so it	
	comfortably sits in the operators hand.	

SCHEDULE – F: ENDOMOTER - ENDODONTIC ELECTRIC MOTOR

Model Quoted :		
S No	Specifications	Compliance by the Bidder
F1	US FDA/European CE Certified	
F2	Simple, flat control panel and LCD screen	
F3	Should include lightweight handpiece with a small head	
F4	Should have Variable torque	
F5	Should be programmable	
F6	Should have Auto-Torque/Reverse	
F7	On / Off button on the motor handpiece	
F8	The handpiece should be Contra-Angle 16:1	
F9	Should operate with or without foot pedal and can be battery operated as well as electrically powered	
F10	Should have Multiple auto-reverse modes	
F11	Should have adjustable programme selections	
F12	Should include Battery, Adapter and Handpiece stand with the equipment	
F13	Should be able to use any rotary system available.	
F14	Demonstration of the Quoted item is must at the predetermined place by the purchaser	

Compliance by S No **Specifications** the Bidder G1 Should be US FDA / European CE Certified and should be Medical grade stainless steel, Autoclavable G2 Products should carry a lifetime warranty against defects in materials and workmanship, and should repair or replace any product that fails as a result of such defect. Medical grade stainless steel G3 G4 Complete set of upper and lower extraction forceps Forceps should have precise, anatomically designed beaks for a sure, effective grip. Should Include all these forceps. Quantity i. **Upper Anterior**, **Premolar and Roots** 3 Each Forceps ii. Upper Molar (Rt & Lt) and Roots 3 Each Forceps iii. Upper bayonet and third molar 2 Each forceps iv. Lower Anterior, Premolar and Roots 3 each Forceps v. Lower Molar and Roots Forceps 3 each vi. Upper (Rt & Lt) and Lower Cow horn 3 each Forceps vii. Lower third Molar Forceps 3 viii. Child pattern Forceps (Pediatric) 1 set G5 **Complete set of elevators** i. Apex Elevators 2 Set ii. Luxators 4 iii.Coupland's Elevator 5 iv. Straight Elevator 5 v. Cryer's Elevator (Left and right) 4 pairs vi. War Wick James Pattern Elevator 4 each (Straight, Left and right) 2 vii. Root Tip Elevator 10 G6 Scalpel handle - Bard Parker style with metric ruler (No 03) G7 **Periosteal Elevators** Molt no.9 25 Howarths 10 **G8** Suture cutting Scissors; Goldman fox 3 each (curve & straight)

SCHEDULE – G: ORAL SURGERY INSTRUMENTS

S No	Specifications		Compliance by the Bidder
G9	Needle Holder, Mayo-Hegar 6"	5	
G10	Adson Plain (Small)	3	
	Adson Toothed (Small)	3	
G11	Bone files Double ended. Cross-Cut file with Curved working end combined with a straight working end Cross-Cut file with Small paddle shaped end is half the size of the larger paddle shaped end.	5	
G12	Bone Curettes – Lucas Type	4	
G13	Bone Rongeurs End cutting	4	
G14	Sponge holder & Dressing forceps (with rachet)	5	
G15	Cheatlle forceps	4	
G16	Wire Cutter	1	
G17	Wire Twister	2	
G18	 Aspirating Syringes Should be made of surgical grade stainless steel. These syringes are indicated for the injection of local anesthetic solution. Ergonomic and lightweight Designed to Reduce hand stress during injection Should accept 1.7 ml and 1.8 ml dental cartridges 	5	

Schedule H: GENERAL DENTAL INSTRUMENTS

S No	Specifications	Quantity	Compliance by the
		Required	Bidder
H1	All the equipment should be US FDA/ I	European CE	
	Certified		
H2	Medical grade stainless steel, Autoclava	able wherever	
	applicable.		
H3	Diagnostics		
	i. Mouth mirror, with Anti Fog,	25 Handles with	
	Mirrors, rust free , slimeline	100 Mirror tops	
	ii. Cordless LED Mouth Mirror	2 handpiece with	
	(Autoclavable) -	10 Mirrors	
	iii. Straight Probe	50	

S No	Specifications	Quantity	Compliance by the
		Required	Bidder
	iv. Dental Explorer – Double	25	
	ended		
	v. Tweezers	25	
H4	Autoclavable, Stainless Steel		
	Instrument tray with lid		
	i. Small - 8 x 6 inch	5	
	ii. Medium - 12 x 8	4	
H5	Kidney trays		
	i. Medium Size – 10 Inch	10	
	ii. Small Size – 8 inch	10	
H6	Lecron carver	5	
H7	 Impression trays for Dentulous arches (upper & lower arches) Should be Autoclavable & reusable Anatomic shapes with good peripheral support. Rigid with ergonomic handle. 	2 sets	
	All SizesPerforated		
H8	 Impression trays for edentulous Arches (upper & lower arches) Should be Autoclavable & reusable Anatomic shapes with good peripheral support. Rigid with ergonomic handle. All Sizes Non Perforated 	2 sets	
H9	Stainless steel bowls Made of medical grade 18:8 Stainless Steel corrosion resistant Should be Autoclavable. About 20 ml	10	
H10	Plasterless Articulator. Any type of dental model should be placed without the application of mounting plaster.	1	
H11	Semi adjustable Articulator (regular 3 point Articulator)	1	
H12	Flexible Rubber mixing Bowls (Small and Medium)	5 each	

S No	Specifications	Quantity	Compliance by the
		Required	Bidder
H13	Plaster mixing spatula	3	
H14	Alginate mixing spatula (Curved)	3	
H15	Wax knife	2	
H16	Wax Spatula	2	
H17	Crown preparation Carbide burs set with box (Kit)	2	
H18	Automatic Crown remover	1	
H19	Spirit Lamp	2	
H20	Fox plane	1	
H21	Dental vibrator	1	
H22	Plaster knife	2	
H23	Dental Instrument Mats, should hold atleast 8 instruments , Non-Slip Soft Material, Autoclavable.	5	
H24	Dental fluoride tray	2	
H25	Rubber Dental Base Former Molds (Large, Medium and Small sizes for both the arches)	1 set	
H26	Wax Carving kit full set	1	

SCHEDULE – I: ORTHODONTIC INSTRUMENTS

S No	Specifications	Quantity	Compliance by the Bidder
I1	All the equipment should be US FDA/ E	European CE	
	Certified		
I2	Adams plier	3	
I3	Universal plier	3	
I4	Straight Cutter long and short	1 each	
	Handle		
I5	Distal End Cutter	1	
I6	Pin And ligature cutter	1	
I7	Heavy wire cutter	1	
I8	Straight Wire plier (Bird beak) Short	1 each	
	and long		
I9	Dellarosa	1	
I10	Bonding Tweezers Curved and	2	
	straight		
I11	Preformed Molar Bands with pre	100	
	welded Buccal tube		

S No	Specifications	Quantity	Compliance by the Bidder
I12	Spot Welder	1	
I13	Solder torch	1	
I14	Chair Side Sand blasting unit	1	
I15	Nola Dry Field System	4	
I16	Intra Oral Mirror set	2	

SCHEDULE – J: ENDODONTIC AND RESTORATIVE INSTRUMENTS

S No	Specifications	Quantity	Compliance by the Bidder
J1	All the equipment should be US FDA/ Europe	ean CE	
	Certified		
J2	ENDODONTIC INSTRUMENTS		
	a. Endodontic Hand Spreaders	2 set	
	b. Endodontic Hand pluggers - Apical	2 set	
	c. Endodontic pluggers – Retro	1	
	Filling d. Endodontic box with Ruler, along with Cleaner box along. Autoclavable. Should accomadate Min 50 finger instruments	5	
	e. Cavity Access Set	1 Set	
	f. Niti K Files All sizes from 08 - 40 with silicone stop marker,	5 boxes each	
	g. Niti K Files All sizes from 45 - 80 with silicone stop marker,	3 Boxes each	
	h. Hedstroems All sizes from 15 – 80 with silicone stop marker	2 boxes each	
	i. Barbed Broach with silicone stop marker	20 Broaches	
	j. Reamer All sizes from 15 – 40 with silicone stop marker	2 boxes each	
	k. Finger Spreader All sizes from 10 – 40	2 boxes each	
	l. Finger plugger All sizes from 10 – 40	2 boxes each	
	m. Lentulo Spiral	2 Set	
	n. Peeso Reamer, Size 1-6.	2 box of each	

S No	Specifications	Quantity	Compliance by the Bidder
		size	
	o. Gates Glidden Drills, Size 1-6,	2 box of each	
		size	
	p. ProTaper Universal Hand	1 boxes each	
	Sx, S1, S2, F1, F2, F3		
	q. ProTaper Universal Rotary	1 box each	
	Sx, S1, S2, F1, F2, F3		
	r. ProTaper Universal Retreatment Kit	2 Set	
	s. Protaper Universal Gutta Percha	5 box each	
	Points F1, F2, F3.	size	
	(Atleast 60 points per set)		
	t. Protaper Universal Paper Points	5 box each	
	F1, F2, F3,		
	(Atleast 50 points per set)		
	u. Fiber Post's translucent, light	01 kits	
	conducting glass fiber resin matrix		
	Basic Kit		
J3	RESTORATIVE INSTRUMENTS		
	Restorative Filling and carving		
	Instruments SetFull Set of instruments. Both anterior and		
	 Put set of instruments. Both anterior and Posterior Restoration. 		
	• Products should include a wide range of		
	 instruments ideally tailored for the purpose. Should be Polished, corrosion-resistant stainless		
	steel tips prevent the adhesion of filling material.		
	• Should not cause corrosion or felting of the		
	filling material Should Be EN CE Approved		
	a. Tofelmere	5	
	b. Sectional Matrix kit	2	
	c. Cement Spatula	5	
	d. Cement Carrier	5	
	e. Cement Condenser	5	
	f. Spoon Excavators	5	
	g. Amalgam carvers Full set	2 each	
		2 1	
	h. Amalgam condensers Full set	2 each	
	h. Amalgam condensers Full set i. Burnishes Full set	2 each	

S	Specifications	Quantity	Compliance by the
No			Bidder
	k. Glass Slab for mixing	2	
J4	Composite Filling Instruments (with	1 set	
	titanium tips)		
J5	Mylar Strips	1000 strips	
J6	Agate spatulas	5	
J7	VITA 3D shade guide	1	
J8	Bur Box that organizes and securely holds	1	
	atleast 20 burs for ultrasonic cleaning		
	and sterilization.		
J9	Composite Polishing kit	1 Kit	
J10	Amalgamator (Non Capsule type)	1	
	• Should not be one meant for amalgam		
	capsules.		

SCHEDULE K: PERIODONTICS KIT

S	Specifications	Quantity	Compliance by the
No			Bidder
K1	All the equipment should be US FDA/ Europe	ean CE	
	Certified		
K2	Scalars		
	i. Anterior Sickle/Hoe Scaler	2	
	ii. Jacquette Scaler anterior and	2 each	
	posterior		
K3	Curettes		
	i. Furcation Curettes	1 set	
	ii. Gracey Curettes	1 Set	
K4	Periodontal Surgery instrument kit (files	1 set	
	and Knives)		
	Should include Orbans, Krikland		
	knives and Slugher and Sugerman		
	files also.		
	• Should have an autoclavable container		
	for the same.		
K5	Implant Maintenance Kit	1 kit	
K6	Williams Periodontal Probe	4	
K7	CPITN probe	2	

MANUFACTURER'S / PRINCIPAL'S AUTHORIZATION FORM

(Clause 13 (c) of the tender)

То

The Administrative Officer, All India Institute of Medical Sciences Raipur

Dear Sir,

TENDER: ______.

we, _who are established reputable manufacturers of and having factories and_ hereby at (name and address of agents) to bid, authorize Messrs. negotiate the contract with against and conclude you Tender goods manufactured No. for the above by individual other us. No company or firm or than Messrs. are authorized to bid, negotiate and conclude the contract in regard to this business against this specific tender.

We hereby extend our full guarantee and warranty as per the conditions of tender for the goods offered for supply against this tender by the above firm.

The authorization is valid up to ______

Yours faithfully,

(Name)

For and on behalf of Messrs. ______ (Name of manufacturers)/Principal.